

gus diseases; common cold; chickenpox; and measles; all childhood diseases; heart disease; heart attacks; essential hypertension; diabetes; insomnia; sterility; difficult labor; morning sickness; nervousness; tension; irritability; itching scalp and skin; numbness; cold hands and feet; dizziness; mental retardation; tooth decay; falling hair; breaking fingernails; paranasal sinusitis; seepage from sinuses; asthma; hay fever; facial neuralgia; retarded growth; pyelitis; thickened blood; ringing in ears; impaired hearing; Meniere's syndrome; callouses and corns; slow healing of cuts and bruises; pimples; tic; cramps in muscles; blocked and swollen lymph glands; cough; infant colic; bed wetting; hangovers; alcoholism; and to provide vigor; promote longevity; maintain good health from the cradle to the grave; to control and reduce weight without restrictions of diet; and to reduce or eliminate the difficulties of old age; and 502(b)(1)—the article failed to bear a label containing the name and place of business of the manufacturer, packer, or distributor.

DISPOSITION: 6-27-60. Default—one bottle and the book "Folk Medicine" were delivered to the Food and Drug Administration and the remainder of the article and its accompanying labeling was destroyed.

6238. Pleasant Aire devices. (F.D.C. No. 44565. S. No. 3-397 R.)

QUANTITY: 86 devices consisting of 67 Model S (room size) devices, 18 Model L dual devices, and 1 Model F furnace device, at Washington, D.C.

SHIPPED: 1-28-60, from Bethesda, Md., by Pleasant Aire, Div. of Viritol Corp.

LABEL IN PART: (Devices) (front) "Pleasant Aire," (back) "Caution Protect Your Eyes!"

ACCOMPANYING LABELING: Leaflets headed "Just Plug in Pleasant Aire" and a poster headed "Lansburgh's First in the Nation to introduce."

RESULTS OF INVESTIGATION: The Model S device was a portable-type cabinet containing a fan, three ultra-violet lamps, and a filter. The dual unit (Model L) contained two blower fans, 5 ultra-violet lamps, and two filters. The furnace unit (Model F) consisted of a transformer and ballast in a metal case and an ultra-violet bactericidal tube to be placed inside the cold air return duct of a furnace.

LIBELED: 5-9-60, Dist. Columbia.

CHARGE: 502(a)—when shipped, the labeling which accompanied the article contained false and misleading representations that the article was an adequate and effective treatment for relieving hay fever, sinus, asthma, and airborne allergies; protecting against many germ and virus produced diseases; preventing spread of contagious airborne germs and viruses; reducing danger of airborne infection; and sanitizing the air.

DISPOSITION: 5-19-60. Consent—claimed by Viritol Corp., and relabeled.

6239. Niblack System spot reducing devices. (F.D.C. No. 44556. S. No. 27-708 R.)

QUANTITY: 10 devices of 3 different styles consisting of 4 Salon Table Units bearing the letters "NS," 5 Salon Pony Units bearing the letters "NS" and 1 Home Pony Unit labeled "The Niblack System Home Pony * * * NS," at Minneapolis, Minn.

SHIPPED: From Denver, Colo., by Niblack System, Inc. The Salon Table Units and Salon Pony Units were shipped some time in 1952, and the Home Pony Unit was shipped on 3-9-60.